

t:slim X2

Insulin Pump

NOW FEATURING
Basal-IQ Technology



Predicts and
helps prevent
lows with zero
fingersticks*

DATE/TIME

LOCATION

CONTACT

*If your glucose alerts and readings from the G6 do not match symptoms or expectations, use a blood glucose meter to make diabetes treatment decisions.



Basal-IQ technology is not a substitute for active self-management of diabetes and does not prevent hypoglycemia in all scenarios. Please visit tandemdiabetes.com/tslimX2-use for more information.



(833) 509-3598
tandemdiabetes.ca

t:simulator App
A free virtual pump demo



Important Safety Information: Caution: Federal (USA) law restricts the t:slim X2 insulin pump and the t:slim X2 insulin pump with Basal-IQ technology to sale by or on the order of a physician. The t:slim X2 insulin pump with Basal-IQ technology (the System) consists of the t:slim X2 insulin pump, which contains Basal-IQ technology, and a compatible integrated continuous glucose monitor (iCGM, sold separately). The t:slim X2 insulin pump is intended for the subcutaneous delivery of insulin, at set and variable rates, for the management of diabetes mellitus in people requiring insulin. The t:slim X2 insulin pump can be used solely for continuous insulin delivery and as part of the System. When used with a compatible iCGM, the t:slim X2 insulin pump with Basal-IQ technology can be used to suspend insulin delivery based on CGM sensor readings. The pump and the System are indicated for use in individuals 6 years of age and greater. The pump and the System are intended for single patient use. The pump and the System are indicated for use with NovoRapid or Humalog U-100 insulin. The System is not indicated for use in pregnant women, people on dialysis, or critically ill patients. Users of the pump and the System must: be willing and able to use the insulin pump, CGM, and all other system components in accordance with their respective instructions for use; test blood glucose levels as recommended by their healthcare provider; demonstrate adequate carb-counting skills; maintain sufficient diabetes self-care skills; see healthcare provider(s) regularly; and have adequate vision and/or hearing to recognize all functions of the pump, including alerts. The t:slim X2 pump, transmitter, and sensor must be removed before MRI, CT, or diathermy treatment. For additional important safety information, visit tandemdiabetes.com/safetyinfo.

© 2020 Tandem Diabetes Care, Inc. All rights reserved. Tandem Diabetes Care, Basal-IQ, and t:slim X2 are either registered trademarks or trademarks of Tandem Diabetes Care, Inc. in the United States and/or other countries. Dexcom and Dexcom G6 are registered trademarks of Dexcom, Inc. in the United States and/or other countries. All other third-party marks are the property of their respective owners. CM-001195_B